
Midmark M11 - v1.0.0
Total Cycles: 3764

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2023 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 72.7 F 0.0
4:00 95.4 F 0.0
6:00 114.5 F 0.1
8:00 142.4 F 1.4
10:00 167.8 F 3.1
12:00 190.2 F 5.6
14:00 204.7 F 9.2
16:00 215.3 F 12.9
18:00 220.7 F 5.6
20:00 230.5 F 9.6
22:00 238.3 F 11.3
24:00 239.3 F 10.6
26:00 235.6 F 8.1
28:00 248.4 F 14.4
30:00 259.1 F 20.2
32:00 268.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.8
2:00 271.6 F 28.4
3:00 272.3 F 28.7
4:00 271.5 F 28.4

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 188.4 F 0.0
10:00 198.2 F 0.0
15:00 201.6 F 0.0
20:00 229.3 F 0.0
25:00 230.1 F 0.0
30:00 227.3 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 33:11
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:11:21
CYCLE COMPLETE

09/13/2023 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____