
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000042
11:04 21/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
CYCLE ABORTED 2:48
WATER LEVEL LOW
NOT STERILE
END TIME 2:48

Digital Signature #
97FC3AA440970F80
