
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008355
8:48 26/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:27
STERILIZING 4:32
136.0 C 220kPa 4:32
136.0 C 217kPa 5:02
136.0 C 220kPa 5:32
136.3 C 223kPa 6:02
136.4 C 222kPa 6:32
136.1 C 220kPa 7:02
136.3 C 220kPa 7:32
136.2 C 221kPa 8:02
136.2 C 221kPa 8:02
VENTING 8:02
AIR DRYING 8:51
CYCLE COMPLETE 10:53

Digital Signature #
DBC393E903899B48
