
Midmark M11 - v1.0.0
Total Cycles: 3038

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2023 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.8 F 0.0
2:00 153.0 F 2.1
4:00 175.6 F 5.9
6:00 197.5 F 10.5
8:00 214.3 F 4.6
10:00 228.3 F 9.9
12:00 231.8 F 8.2
14:00 241.5 F 12.2
16:00 239.8 F 10.4
18:00 253.3 F 17.1
20:00 264.9 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 29.1
2:00 271.5 F 28.5
3:00 272.6 F 29.2
4:00 271.7 F 28.5

Min 271.4 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 209.8 F 0.0
10:00 230.4 F 0.0
15:00 223.1 F 0.0
20:00 221.0 F 0.0
25:00 221.1 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:03
CYCLE COMPLETE

03/27/2023 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____