
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000250
9:49 21/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:37
STERILIZING 5:00
136.5 C 223kPa 5:00
136.4 C 224kPa 5:30
136.3 C 224kPa 6:00
136.0 C 221kPa 6:30
136.4 C 220kPa 7:00
136.3 C 221kPa 7:30
136.4 C 224kPa 8:00
136.3 C 220kPa 8:30
136.3 C 220kPa 8:30
VENTING 8:30
AIR DRYING 9:25
CYCLE COMPLETE 9:27

Digital Signature #
52A227A10A1D2C4B
