

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5465

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2025 03:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.0 F 0.0  
2:00 141.1 F 2.1  
4:00 156.0 F 5.4  
6:00 184.0 F 9.2  
8:00 203.2 F 13.3  
10:00 215.1 F 4.9  
12:00 225.0 F 8.8  
14:00 235.3 F 12.8  
16:00 235.6 F 9.0  
18:00 242.2 F 11.5  
20:00 243.2 F 12.0  
22:00 254.3 F 17.7  
24:00 264.3 F 23.7  
26:00 272.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 271.0 F 28.6  
2:00 272.3 F 29.6  
3:00 272.5 F 29.6  
4:00 272.1 F 29.1

Min 270.9 F 28.4  
Max 273.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 198.1 F 0.0  
10:00 194.7 F 0.0  
15:00 243.3 F 0.0  
20:00 230.6 F 0.0  
25:00 229.2 F 0.0  
30:00 224.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 26:10  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:04:56  
CYCLE COMPLETE

03/31/2025 04:25 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_