
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002344
11:10 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:11
STERILIZING 4:20
136.0 C 223kPa 4:20
136.3 C 226kPa 4:50
136.4 C 226kPa 5:20
136.4 C 224kPa 5:50
136.2 C 224kPa 6:20
136.4 C 226kPa 6:50
136.2 C 225kPa 7:20
136.4 C 227kPa 7:50
136.4 C 227kPa 7:50
VENTING 7:50
AIR DRYING 8:44
CYCLE COMPLETE 9:22

Digital Signature #
B634AA7CC3B8526C
