

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3650

Name: RITTER 1  
Sterilizer ID: V2178198  
08/14/2023 07:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.3 F 0.0  
2:00 164.0 F 3.6  
4:00 185.0 F 8.3  
6:00 210.3 F 9.0  
8:00 225.6 F 9.1  
10:00 233.5 F 8.9  
12:00 241.9 F 12.5  
14:00 241.5 F 11.2  
16:00 255.3 F 18.2  
18:00 267.3 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 271.7 F 28.6  
2:00 271.9 F 28.8  
3:00 271.9 F 28.8  
4:00 271.9 F 28.8  
  
Min 271.5 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.1 F 0.0  
5:00 224.8 F 0.0  
10:00 220.4 F 0.0  
15:00 225.8 F 0.0  
20:00 230.1 F 0.0  
25:00 224.8 F 0.0  
30:00 223.3 F 0.0  
  
Drying Complete

FILLING: 1:37  
HEATING: 19:13  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:03  
CYCLE COMPLETE

08/14/2023 08:34 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_