
Midmark M11 - v1.0.5
Total Cycles: 2553

Name: RITTER2
Sterilizer ID: V2313228
06/22/2023 12:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.6 F 0.0
2:00 164.0 F 2.7
4:00 175.5 F 6.8
6:00 200.3 F 11.8
8:00 217.6 F 6.4
10:00 231.8 F 11.7
12:00 233.2 F 8.8
14:00 239.1 F 10.2
16:00 246.6 F 13.9
18:00 258.8 F 20.6
20:00 269.8 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.8
2:00 272.3 F 29.3
3:00 271.3 F 28.8
4:00 271.8 F 29.5

Min 271.2 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 216.7 F 0.0
10:00 216.9 F 0.0
15:00 218.0 F 0.0
20:00 216.6 F 0.0
25:00 214.9 F 0.0
30:00 213.5 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:00
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

06/22/2023 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____