

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001651  
15:28 27/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:44  
STERILIZING 3:44  
136.0 C 220kPa 3:44  
136.4 C 226kPa 4:14  
136.4 C 225kPa 4:44  
136.4 C 225kPa 5:14  
136.4 C 224kPa 5:44  
136.4 C 226kPa 6:14  
136.5 C 223kPa 6:44  
136.4 C 222kPa 7:14  
136.4 C 222kPa 7:14  
VENTING 7:14  
AIR DRYING 8:13  
CYCLE COMPLETE 8:25

-----  
Digital Signature #  
DCA7EF74FF221686  
-----