
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004156
8:34 08/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:31
STERILIZING 3:41
136.4 C 222kPa 3:41
136.0 C 222kPa 4:11
136.3 C 223kPa 4:41
136.4 C 223kPa 5:11
136.1 C 219kPa 5:41
136.3 C 223kPa 6:11
136.3 C 223kPa 6:41
136.3 C 221kPa 7:11
136.3 C 221kPa 7:11
VENTING 7:11
AIR DRYING 7:58
CYCLE COMPLETE 8:30

Digital Signature #
17C3FB01C4C60D6F
