
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007171
7:24 09/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:40
STERILIZING 4:35
136.0 C 220kPa 4:35
136.1 C 222kPa 5:05
136.4 C 223kPa 5:35
136.4 C 220kPa 6:05
136.2 C 222kPa 6:35
136.3 C 222kPa 7:05
136.2 C 221kPa 7:35
136.4 C 221kPa 8:05
136.4 C 221kPa 8:05
VENTING 8:05
AIR DRYING 8:55
CYCLE COMPLETE 9:03

Digital Signature #
3466DEBC574AAEB1
