

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1497

Name: RITTER2  
Sterilizer ID: V2313228  
09/15/2022 11:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
2:00 179.0 F 3.8  
4:00 188.9 F 8.6  
6:00 209.8 F 8.1  
8:00 226.9 F 9.3  
10:00 235.1 F 9.8  
12:00 242.2 F 12.7  
14:00 242.3 F 11.9  
16:00 256.2 F 19.1  
18:00 268.2 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 272.6 F 29.5  
2:00 271.8 F 29.1  
3:00 272.7 F 29.5  
4:00 272.0 F 28.9  
  
Min 271.8 F 28.8  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 217.8 F 0.0  
10:00 228.0 F 0.0  
15:00 227.9 F 0.0  
20:00 223.5 F 0.0  
25:00 221.2 F 0.0  
30:00 218.6 F 0.0  
Drying Complete

FILLING: 2:38  
HEATING: 19:12  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:59  
CYCLE COMPLETE

09/15/2022 12:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_