

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2628

Name: RITTER2  
Sterilizer ID: V2313228  
07/10/2023 12:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
2:00 182.0 F 3.9  
4:00 193.7 F 8.5  
6:00 213.4 F 10.9  
8:00 226.2 F 8.5  
10:00 237.7 F 12.1  
12:00 239.9 F 11.6  
14:00 238.9 F 10.4  
16:00 252.4 F 17.1  
18:00 264.3 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.7 F 29.7  
2:00 271.5 F 29.1  
3:00 272.9 F 29.8  
4:00 271.7 F 29.0  
  
Min 271.5 F 28.9  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
5:00 222.5 F 0.0  
10:00 228.1 F 0.0  
15:00 228.6 F 0.0  
20:00 229.6 F 0.0  
25:00 228.8 F 0.0  
30:00 227.0 F 0.0  
  
Drying Complete

FILLING: 1:53  
HEATING: 19:55  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:03  
CYCLE COMPLETE

07/10/2023 01:43 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_