
Midmark M11 - v1.0.0
Total Cycles: 4729

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2024 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.6 F 0.0
2:00 129.9 F 1.8
4:00 150.3 F 4.7
6:00 175.3 F 8.3
8:00 200.2 F 12.4
10:00 216.1 F 4.1
12:00 224.8 F 7.9
14:00 235.8 F 11.9
16:00 233.1 F 7.9
18:00 244.6 F 13.1
20:00 242.5 F 11.4
22:00 254.2 F 17.4
24:00 264.6 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.2 F 28.9
2:00 272.0 F 28.5
3:00 271.1 F 28.4
4:00 272.8 F 29.3

Min 271.0 F 28.1
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 228.1 F 0.0
10:00 212.2 F 0.0
15:00 208.8 F 0.0
20:00 208.1 F 0.0
25:00 221.4 F 0.0
30:00 188.3 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 25:53
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:04:50
CYCLE COMPLETE

07/15/2024 02:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____