
Midmark M11 - v1.0.5
Total Cycles: 993

Name: RITTER2
Sterilizer ID: V2313228
04/14/2022 04:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.4 F 0.0
2:00 191.9 F 3.7
4:00 208.2 F 10.0
6:00 224.3 F 6.9
8:00 237.4 F 12.9
10:00 240.1 F 11.7
12:00 241.2 F 11.4
14:00 255.3 F 18.7
16:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.6
2:00 271.4 F 29.2
3:00 272.6 F 29.8
4:00 271.8 F 29.1

Min 271.3 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 230.0 F 0.0
10:00 229.7 F 0.0
15:00 228.9 F 0.0
20:00 229.3 F 0.0
25:00 229.7 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 17:18
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:56:32
CYCLE COMPLETE

04/14/2022 05:05 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____