

-----  
Midmark M11 - v1.0.5  
Total Cycles: 206

Name: RITTER2  
Sterilizer ID: V2313228  
07/14/2021 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.3 F 0.0  
2:00 177.8 F 2.8  
4:00 191.5 F 7.5  
6:00 213.2 F 12.8  
8:00 222.3 F 8.3  
10:00 237.2 F 13.1  
12:00 239.4 F 11.7  
14:00 239.6 F 10.7  
16:00 252.6 F 17.2  
18:00 263.7 F 23.8  
20:00 272.0 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.1  
1:00 272.5 F 29.9  
2:00 271.6 F 29.0  
3:00 272.5 F 29.8  
4:00 271.8 F 29.1  
  
Min 271.4 F 29.0  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.1 F 0.0  
5:00 227.3 F 0.0  
10:00 223.4 F 0.0  
15:00 223.8 F 0.0  
20:00 224.2 F 0.0  
25:00 226.7 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 20:01  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:13  
CYCLE COMPLETE

07/14/2021 10:54 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01