
Midmark M11 - v1.0.5
Total Cycles: 3507

Name: RITTER2
Sterilizer ID: V2313228
04/09/2024 08:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 70.7 F 0.0
4:00 77.9 F 0.0
6:00 107.9 F 0.1
8:00 158.0 F 0.4
10:00 182.0 F 2.9
12:00 199.8 F 6.3
14:00 218.3 F 9.8
16:00 226.3 F 12.3
18:00 228.1 F 7.9
20:00 238.1 F 12.2
22:00 234.6 F 9.0
24:00 241.0 F 10.8
26:00 245.5 F 13.3
28:00 257.1 F 19.5
30:00 267.5 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.4 F 29.5
2:00 272.0 F 28.8
3:00 272.3 F 29.5
4:00 272.3 F 29.1

Min 271.4 F 28.6
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 204.8 F 0.0
10:00 199.9 F 0.0
15:00 193.2 F 0.0
20:00 194.3 F 0.0
25:00 197.2 F 0.0
30:00 203.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:18
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

04/09/2024 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____