
Midmark M11 - v1.0.0
Total Cycles: 1752

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2022 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.9 F 0.0
2:00 180.9 F 2.8
4:00 184.0 F 6.6
6:00 206.2 F 11.6
8:00 224.9 F 6.1
10:00 236.5 F 11.4
12:00 234.5 F 8.3
14:00 241.0 F 10.8
16:00 246.1 F 13.2
18:00 258.7 F 19.9
20:00 270.0 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.6 F 29.2
2:00 272.3 F 28.8
3:00 271.7 F 28.8
4:00 272.8 F 29.1

Min 271.2 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 235.5 F 0.0
10:00 200.0 F 0.0
15:00 199.0 F 0.0
20:00 203.7 F 0.0
25:00 215.2 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

03/28/2022 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____