
Midmark M11 - v1.0.5
Total Cycles: 318

Name: RITTER2
Sterilizer ID: V2313228
08/30/2021 08:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.2 F 0.0
2:00 127.0 F 1.1
4:00 144.5 F 4.3
6:00 172.4 F 8.6
8:00 196.5 F 12.1
10:00 214.7 F 6.2
12:00 228.4 F 11.0
14:00 231.6 F 7.8
16:00 243.5 F 12.5
18:00 243.5 F 12.7
20:00 256.3 F 19.3
22:00 267.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.2
1:00 272.7 F 29.9
2:00 271.9 F 29.0
3:00 272.4 F 29.9
4:00 272.2 F 29.2

Min 271.5 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.8 F 0.0
5:00 247.3 F 0.0
10:00 207.8 F 0.0
15:00 192.4 F 0.0
20:00 179.9 F 0.0
25:00 182.0 F 0.0
30:00 177.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

08/30/2021 09:15 AM
ACCEPTED: X
REJECTED:
Operator: U02