
Midmark M11 - v1.0.5
Total Cycles: 2128

Name: RITTER2
Sterilizer ID: V2313228
03/08/2023 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 194.5 F 4.3
4:00 201.8 F 9.3
6:00 222.3 F 7.1
8:00 231.2 F 10.7
10:00 232.3 F 8.3
12:00 239.4 F 10.3
14:00 247.7 F 14.5
16:00 260.4 F 21.6
18:00 271.9 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.4 F 29.4
2:00 271.8 F 29.4
3:00 272.2 F 29.3
4:00 271.4 F 29.2

Min 271.3 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.5 F 0.0
5:00 222.7 F 0.0
10:00 241.8 F 0.0
15:00 242.4 F 0.0
20:00 234.7 F 0.0
25:00 224.5 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:42
CYCLE COMPLETE

03/08/2023 03:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____