
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006438
8:34 11/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:00
STERILIZING 4:06
136.5 C 220kPa 4:06
136.2 C 222kPa 4:36
136.3 C 221kPa 5:06
136.5 C 221kPa 5:36
136.1 C 220kPa 6:06
136.4 C 222kPa 6:36
136.2 C 219kPa 7:06
136.3 C 220kPa 7:36
136.3 C 220kPa 7:36
VENTING 7:37
AIR DRYING 8:26
CYCLE COMPLETE 8:33

Digital Signature #
FBF04A40D250AD39
