
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001614
15:34 11/24/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:14
PRESSURIZING 5:54
STERILIZING 7:12
136.0 C 220kPa 7:12
135.9 C 218kPa 7:42
136.2 C 221kPa 8:12
136.4 C 223kPa 8:42
136.2 C 221kPa 9:12
136.5 C 223kPa 9:42
136.4 C 223kPa 10:12
136.3 C 221kPa 10:42
VENTING 10:42
AIR DRYING 11:42
CYCLE COMPLETE 41:42

Digital Signature #
6D11E2817C79F0A3
