
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004906
17:10 04/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:30
STERILIZING 3:27
136.4 C 224kPa 3:27
136.3 C 224kPa 3:57
136.3 C 224kPa 4:27
136.1 C 220kPa 4:57
136.4 C 225kPa 5:27
136.4 C 222kPa 5:57
136.4 C 225kPa 6:27
136.4 C 224kPa 6:57
136.4 C 224kPa 6:57
VENTING 6:57
AIR DRYING 7:42
CYCLE COMPLETE 9:30

Digital Signature #
E22D1E812650AE72
