
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002420
10:16 14/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:03
STERILIZING 4:13
136.0 C 221kPa 4:13
136.3 C 224kPa 4:43
136.4 C 225kPa 5:13
136.3 C 223kPa 5:43
136.3 C 223kPa 6:13
136.4 C 223kPa 6:43
136.4 C 222kPa 7:13
136.2 C 222kPa 7:43
136.2 C 222kPa 7:43
VENTING 7:43
AIR DRYING 8:37
CYCLE COMPLETE 18:35

Digital Signature #
679F5CA82D224741
