
Midmark M11 - v1.0.0
Total Cycles: 1701

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2022 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.4 F 0.0
2:00 193.4 F 3.7
4:00 196.4 F 8.1
6:00 220.7 F 12.3
8:00 229.3 F 8.7
10:00 239.1 F 10.7
12:00 240.0 F 11.6
14:00 238.9 F 10.0
16:00 252.4 F 16.7
18:00 264.2 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.3 F 28.9
2:00 272.3 F 29.3
3:00 271.5 F 28.6
4:00 272.4 F 29.2

Min 271.4 F 28.5
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 232.7 F 0.0
10:00 231.8 F 0.0
15:00 229.6 F 0.0
20:00 228.1 F 0.0
25:00 229.2 F 0.0
30:00 231.9 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

03/09/2022 12:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____