

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3176

Name: RITTER 1  
Sterilizer ID: V2178198  
04/28/2023 10:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.8 F 0.0  
2:00 190.6 F 2.3  
4:00 199.4 F 6.9  
6:00 216.8 F 11.8  
8:00 224.3 F 6.7  
10:00 236.4 F 11.9  
12:00 236.4 F 9.3  
14:00 239.2 F 9.6  
16:00 248.5 F 14.3  
18:00 260.4 F 20.9  
20:00 271.1 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.4 F 28.7  
2:00 271.7 F 28.5  
3:00 272.4 F 28.7  
4:00 271.5 F 28.2

Min 271.5 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 214.3 F 0.0  
10:00 220.2 F 0.0  
15:00 221.5 F 0.0  
20:00 221.5 F 0.0  
25:00 222.5 F 0.0  
30:00 219.8 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 20:42  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:59:59  
CYCLE COMPLETE

04/28/2023 11:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_