

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005951  
8:07 12/07/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:28  
STERILIZING 3:43  
136.4 C 223kPa 3:43  
136.3 C 223kPa 4:13  
136.3 C 219kPa 4:43  
136.3 C 222kPa 5:13  
136.3 C 222kPa 5:43  
136.2 C 220kPa 6:13  
136.4 C 224kPa 6:43  
136.4 C 219kPa 7:13  
136.4 C 219kPa 7:13  
VENTING 7:13  
AIR DRYING 8:05  
CYCLE COMPLETE 14:50

-----  
Digital Signature #  
40A4A9479393621A  
-----