
Midmark M11 - v1.0.5
Total Cycles: 3066

Name: RITTER2
Sterilizer ID: V2313228
10/30/2023 03:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.8 F 0.0
2:00 136.2 F 1.1
4:00 162.9 F 4.0
6:00 183.2 F 7.9
8:00 203.2 F 12.3
10:00 219.4 F 6.3
12:00 229.3 F 11.2
14:00 233.0 F 8.2
16:00 240.0 F 11.4
18:00 245.7 F 13.4
20:00 257.8 F 20.0
22:00 268.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.2 F 29.4
2:00 272.8 F 29.8
3:00 272.3 F 30.0
4:00 271.2 F 29.3

Min 271.2 F 29.3
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.1 F 0.0
5:00 229.7 F 0.0
10:00 219.4 F 0.0
15:00 190.1 F 0.0
20:00 181.4 F 0.0
25:00 179.5 F 0.0
30:00 177.7 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:47
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:27
CYCLE COMPLETE

10/30/2023 04:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____