
Midmark M11 - v1.0.0
Total Cycles: 5235

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2025 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.8 F 0.0
2:00 110.6 F 0.0
4:00 133.7 F 0.0
6:00 172.4 F 0.5
8:00 188.4 F 4.0
10:00 199.7 F 7.8
12:00 215.0 F 11.7
14:00 224.0 F 5.8
16:00 235.9 F 10.6
18:00 233.4 F 7.8
20:00 245.4 F 13.1
22:00 243.0 F 11.7
24:00 255.0 F 17.7
26:00 265.4 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.1
1:00 271.8 F 28.3
2:00 272.0 F 28.4
3:00 272.3 F 28.7
4:00 272.1 F 28.8

Min 271.4 F 28.0
Max 272.4 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 231.0 F 0.0
10:00 223.5 F 0.0
15:00 220.8 F 0.0
20:00 234.2 F 0.0
25:00 229.3 F 0.0
30:00 227.3 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 27:41
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:06:21
CYCLE COMPLETE

01/10/2025 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____