
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000993
10:11 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:38
STERILIZING 3:43
136.4 C 233kPa 3:43
136.3 C 232kPa 4:13
136.2 C 231kPa 4:43
136.4 C 233kPa 5:13
136.4 C 233kPa 5:43
136.4 C 231kPa 6:13
136.4 C 233kPa 6:43
136.4 C 233kPa 7:13
136.4 C 233kPa 7:13
VENTING 7:13
AIR DRYING 8:05
CYCLE COMPLETE 8:33

Digital Signature #
75D9DB3B8ACEABB1
