

-----  
Midmark M11 - v1.0.5  
Total Cycles: 656

Name: RITTER2  
Sterilizer ID: V2313228  
12/21/2021 08:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.3 F 0.0  
2:00 76.0 F 0.0  
4:00 109.8 F 0.1  
6:00 147.9 F 0.2  
8:00 169.6 F 0.4  
10:00 188.7 F 3.5  
12:00 207.9 F 7.0  
14:00 218.9 F 10.5  
16:00 228.8 F 8.1  
18:00 230.4 F 9.1  
20:00 240.0 F 13.6  
22:00 238.8 F 11.0  
24:00 236.5 F 9.1  
26:00 249.4 F 15.4  
28:00 260.2 F 21.5  
30:00 270.2 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.4 F 29.5  
2:00 271.8 F 29.5  
3:00 272.1 F 29.3  
4:00 271.9 F 29.6

Min 271.5 F 28.9  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 200.3 F 0.0  
10:00 204.4 F 0.0  
15:00 215.1 F 0.0  
20:00 221.2 F 0.0  
25:00 228.8 F 0.0  
30:00 222.7 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 30:55  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:09:17  
CYCLE COMPLETE

12/21/2021 09:19 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_