
Midmark M11 - v1.0.0
Total Cycles: 820

Name: RITTER 1
Sterilizer ID: V2178198
05/19/2021 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.8 F 0.0
2:00 191.3 F 2.1
4:00 194.9 F 6.4
6:00 213.4 F 12.0
8:00 224.1 F 6.4
10:00 238.7 F 12.2
12:00 238.0 F 9.7
14:00 237.6 F 8.6
16:00 250.6 F 15.3
18:00 263.1 F 22.5
20:00 271.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.2 F 28.3
2:00 272.6 F 28.8
3:00 271.6 F 28.0
4:00 272.5 F 28.6

Min 271.6 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 223.4 F 0.0
10:00 233.5 F 0.0
15:00 224.1 F 0.0
20:00 193.9 F 0.0
Dry Cycle Aborted

FILLING: 1:36
HEATING: 20:05
STERILIZING: 4:00
VENTING: 3:03
DRYING: 24:18
TOTAL CYCLE: 00:53:03

05/19/2021 09:59 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____