

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000340  
8:43 15/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:49  
PRESSURIZING 2:56  
STERILIZING 4:10  
136.0 C 228kPa 4:10  
136.4 C 233kPa 4:40  
136.3 C 232kPa 5:10  
136.5 C 233kPa 5:40  
136.4 C 232kPa 6:10  
136.3 C 232kPa 6:40  
136.3 C 232kPa 7:10  
136.3 C 231kPa 7:40  
136.3 C 231kPa 7:40  
VENTING 7:40  
AIR DRYING 8:34  
CYCLE COMPLETE 23:34  
-----

Digital Signature #  
31778CACA1A085A8  
-----