
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000131
9:36 12/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:09
STERILIZING 4:17
136.4 C 225kPa 4:17
136.2 C 222kPa 4:47
136.4 C 224kPa 5:17
136.3 C 223kPa 5:47
136.4 C 223kPa 6:17
135.9 C 219kPa 6:47
136.3 C 223kPa 7:17
136.4 C 222kPa 7:47
136.4 C 222kPa 7:47
VENTING 7:47
AIR DRYING 8:39
CYCLE COMPLETE 8:43

Digital Signature #
61AF14C87FF839F5
