
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 000848
12:01 20/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:39
STERILIZING 4:40
136.4 C 223kPa 4:40
136.4 C 225kPa 5:10
136.3 C 224kPa 5:40
136.4 C 222kPa 6:10
136.2 C 223kPa 6:40
136.4 C 225kPa 7:10
136.2 C 221kPa 7:40
136.4 C 222kPa 8:10
136.4 C 222kPa 8:10
VENTING 8:10
AIR DRYING 9:05
CYCLE COMPLETE 9:21

Digital Signature #
2E292B1ED195C493
