

-----  
Midmark M11 - v1.0.5  
Total Cycles: 963

Name: RITTER2  
Sterilizer ID: V2313228  
04/06/2022 01:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.5 F 0.0  
2:00 132.8 F 0.0  
4:00 156.0 F 0.5  
6:00 185.2 F 4.0  
8:00 204.2 F 8.3  
10:00 221.4 F 13.2  
12:00 229.7 F 9.0  
14:00 238.4 F 10.9  
16:00 242.8 F 12.7  
18:00 242.9 F 12.1  
20:00 256.4 F 19.1  
22:00 268.1 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.5 F 29.4  
2:00 271.2 F 28.8  
3:00 272.6 F 29.5  
4:00 271.6 F 28.8

Min 271.2 F 28.7  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
5:00 199.1 F 0.0  
10:00 210.8 F 0.0  
15:00 205.5 F 0.0  
20:00 210.5 F 0.0  
25:00 221.6 F 0.0  
30:00 236.1 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 23:14  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:21  
CYCLE COMPLETE

04/06/2022 02:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_