
Midmark M11 - v1.0.5
Total Cycles: 4110

Name: RITTER2
Sterilizer ID: V2313228
11/12/2024 03:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 83.9 F 0.0
2:00 85.4 F 0.0
4:00 103.2 F 0.0
6:00 135.5 F 0.3
8:00 168.6 F 1.9
10:00 189.8 F 4.9
12:00 203.0 F 8.2
14:00 215.6 F 11.6
16:00 221.3 F 5.8
18:00 232.7 F 10.2
20:00 235.3 F 9.1
22:00 243.3 F 12.8
24:00 242.5 F 11.7
26:00 254.5 F 17.9
28:00 265.2 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 29.0
2:00 272.5 F 29.5
3:00 271.8 F 28.6
4:00 272.8 F 29.3

Min 271.5 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 219.0 F 0.0
10:00 202.7 F 0.0
15:00 197.5 F 0.0
20:00 194.7 F 0.0
25:00 204.3 F 0.0
30:00 207.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:51
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:08:20
CYCLE COMPLETE

11/12/2024 04:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____