

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1152

Name: RITTER 1  
Sterilizer ID: V2178198  
09/08/2021 02:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.0 F 0.0  
2:00 202.4 F 5.2  
4:00 223.4 F 11.5  
6:00 230.1 F 8.5  
8:00 238.1 F 9.9  
10:00 243.6 F 12.2  
12:00 246.3 F 13.6  
14:00 259.6 F 20.8  
16:00 271.1 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.1 F 28.9  
2:00 271.5 F 28.8  
3:00 272.1 F 28.9  
4:00 271.2 F 28.5  
  
Min 271.1 F 28.4  
Max 272.4 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
5:00 230.1 F 0.0  
10:00 231.7 F 0.0  
15:00 231.9 F 0.0  
20:00 220.2 F 0.0  
25:00 209.6 F 0.0  
30:00 193.9 F 0.0  
Drying Complete

FILLING: 3:11  
HEATING: 16:41  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:01  
CYCLE COMPLETE

09/08/2021 03:06 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01