
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000782
9:22 29/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:10
STERILIZING 3:11
136.0 C 228kPa 3:11
136.1 C 232kPa 3:41
136.3 C 233kPa 4:11
136.4 C 233kPa 4:41
136.4 C 231kPa 5:11
136.2 C 232kPa 5:41
136.3 C 233kPa 6:11
136.4 C 233kPa 6:41
136.4 C 233kPa 6:41
VENTING 6:41
AIR DRYING 7:32
CYCLE COMPLETE 7:33

Digital Signature #
DFD8CC5335F3D68B
