
Midmark M11 - v1.0.5
Total Cycles: 1460

Name: RITTER2
Sterilizer ID: V2313228
09/07/2022 08:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.7 F 0.0
2:00 76.1 F 0.0
4:00 92.0 F 0.0
6:00 111.2 F 0.0
8:00 137.3 F 1.2
10:00 169.7 F 2.9
12:00 194.8 F 5.0
14:00 212.9 F 8.1
16:00 224.8 F 11.7
18:00 224.5 F 6.5
20:00 235.7 F 11.3
22:00 233.3 F 8.2
24:00 242.7 F 12.0
26:00 244.0 F 12.6
28:00 255.6 F 18.8
30:00 266.2 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.4 F 29.2
2:00 271.9 F 29.4
3:00 272.0 F 28.9
4:00 272.3 F 29.6

Min 271.4 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 200.9 F 0.0
10:00 212.1 F 0.0
15:00 232.0 F 0.0
20:00 224.9 F 0.0
25:00 225.7 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:41
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:09:34
CYCLE COMPLETE

09/07/2022 09:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____