
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007788
11:20 22/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:29
STERILIZING 3:30
136.4 C 224kPa 3:30
136.1 C 222kPa 4:00
136.4 C 223kPa 4:30
136.4 C 222kPa 5:00
136.4 C 223kPa 5:30
136.4 C 224kPa 6:00
136.4 C 225kPa 6:30
136.4 C 223kPa 7:00
136.4 C 223kPa 7:00
VENTING 7:00
AIR DRYING 7:51
CYCLE COMPLETE 16:08

Digital Signature #
CA1AE67C6A1B0D10
