
Midmark M11 - v1.0.5
Total Cycles: 2152

Name: RITTER2
Sterilizer ID: V2313228
03/15/2023 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.8 F 0.0
2:00 114.3 F 0.1
4:00 154.8 F 2.5
6:00 165.2 F 6.8
8:00 191.1 F 11.6
10:00 211.8 F 4.4
12:00 225.7 F 8.9
14:00 236.1 F 11.6
16:00 239.5 F 11.1
18:00 237.5 F 9.5
20:00 251.8 F 16.3
22:00 263.7 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.7 F 29.2
2:00 272.3 F 29.2
3:00 271.8 F 28.5
4:00 272.7 F 29.2

Min 271.5 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.1 F 0.0
5:00 208.1 F 0.0
10:00 211.6 F 0.0
15:00 211.6 F 0.0
20:00 207.2 F 0.0
25:00 205.5 F 0.0
30:00 205.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:51
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:01:48
CYCLE COMPLETE

03/15/2023 03:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____