
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001190
12:28 27/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:48
STERILIZING 5:05
136.2 C 221kPa 5:05
136.3 C 224kPa 5:35
136.1 C 224kPa 6:05
136.4 C 225kPa 6:35
136.4 C 224kPa 7:05
136.4 C 225kPa 7:35
136.3 C 224kPa 8:05
136.3 C 224kPa 8:35
136.3 C 224kPa 8:35
VENTING 8:35
AIR DRYING 9:29
CYCLE COMPLETE 24:29

Digital Signature #
344EA92EB7874D8C
