
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000336
11:56 08/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:45
STERILIZING 4:54
136.0 C 230kPa 4:54
136.3 C 232kPa 5:25
136.4 C 233kPa 5:55
CYCLE ABORTED 6:12
STOP BUTTON PRESSED
NOT STERILE
END TIME 6:12

Digital Signature #
4D9DF0C4D7CF5C58
