

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4126

Name: RITTER2  
Sterilizer ID: V2313228  
11/15/2024 12:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.4 F 0.0  
2:00 147.4 F 0.2  
4:00 171.9 F 3.0  
6:00 198.1 F 6.7  
8:00 214.7 F 10.8  
10:00 222.5 F 6.2  
12:00 230.6 F 10.2  
14:00 235.0 F 9.0  
16:00 241.8 F 12.8  
18:00 241.5 F 11.7  
20:00 253.6 F 18.0  
22:00 265.0 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.2  
1:00 272.3 F 29.5  
2:00 272.0 F 29.9  
3:00 272.7 F 30.7  
4:00 272.3 F 30.0

Min 270.9 F 29.1  
Max 273.6 F 30.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.9 F 0.0  
5:00 217.1 F 0.0  
10:00 218.4 F 0.0  
15:00 206.0 F 0.0  
20:00 199.2 F 0.0  
25:00 194.6 F 0.0  
30:00 195.4 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 23:54  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:02:42  
CYCLE COMPLETE

11/15/2024 01:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_