
Midmark M11 - v1.0.0
Total Cycles: 2366

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2022 07:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.6 F 0.0
2:00 78.9 F 0.0
4:00 96.4 F 0.0
6:00 128.2 F 0.1
8:00 152.0 F 0.2
10:00 179.6 F 2.2
12:00 199.7 F 5.4
14:00 212.1 F 8.6
16:00 223.5 F 12.1
18:00 223.7 F 6.6
20:00 235.2 F 11.3
22:00 233.1 F 8.0
24:00 243.5 F 12.0
26:00 244.4 F 12.6
28:00 256.4 F 19.0
30:00 267.5 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.7 F 29.2
2:00 271.7 F 28.4
3:00 272.4 F 29.3
4:00 272.2 F 28.8

Min 271.4 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.0 F 0.0
5:00 200.6 F 0.0
10:00 210.0 F 0.0
15:00 209.2 F 0.0
20:00 209.2 F 0.0
25:00 209.1 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:27
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

09/26/2022 08:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____