
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000649
7:20 13/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:05
PRESSURIZING 1:14
STERILIZING 2:08
136.0 C 225kPa 2:08
136.4 C 230kPa 2:38
136.4 C 232kPa 3:08
136.3 C 231kPa 3:38
136.4 C 233kPa 4:08
136.1 C 229kPa 4:38
136.4 C 233kPa 5:08
136.4 C 231kPa 5:38
136.4 C 231kPa 5:38
VENTING 5:38
AIR DRYING 6:30
CYCLE COMPLETE 6:31

Digital Signature #
12DC6FEE37CE9FFA
