
Midmark M11 - v1.0.5
Total Cycles: 4410

Name: RITTER2
Sterilizer ID: V2313228
03/07/2025 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 72.1 F 0.2
4:00 77.8 F 0.9
6:00 104.0 F 3.1
8:00 142.8 F 6.1
10:00 167.2 F 9.7
12:00 187.4 F 13.1
14:00 207.1 F 4.3
16:00 218.2 F 7.7
18:00 229.0 F 11.6
20:00 229.0 F 7.5
22:00 241.1 F 12.8
24:00 239.9 F 11.1
26:00 252.1 F 17.2
28:00 262.9 F 23.5
30:00 272.9 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.9 F 30.0
2:00 271.7 F 29.3
3:00 273.0 F 30.1
4:00 272.1 F 29.3

Min 271.6 F 29.1
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.6 F 0.0
5:00 179.9 F 0.0
10:00 194.9 F 0.0
15:00 196.9 F 0.0
20:00 202.4 F 0.0
25:00 208.6 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:19
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:08:41
CYCLE COMPLETE

03/07/2025 11:21 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____