
Midmark M11 - v1.0.0
Total Cycles: 4306

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2024 02:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.2 F 0.0
2:00 89.2 F 0.6
4:00 103.5 F 2.2
6:00 132.3 F 4.5
8:00 159.1 F 7.1
10:00 178.4 F 10.1
12:00 191.8 F 13.2
14:00 205.7 F 4.7
16:00 215.7 F 8.0
18:00 225.1 F 11.6
20:00 226.5 F 6.3
22:00 237.8 F 11.0
24:00 233.8 F 7.8
26:00 245.5 F 13.2
28:00 255.5 F 18.5
30:00 264.7 F 24.0
32:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.5 F 29.1
2:00 272.5 F 29.4
3:00 271.5 F 29.0
4:00 272.4 F 29.1

Min 271.5 F 28.8
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 242.8 F 0.0
10:00 228.0 F 0.0
15:00 229.8 F 0.0
20:00 226.7 F 0.0
25:00 229.0 F 0.0
30:00 227.8 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 32:52
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:11:32
CYCLE COMPLETE

02/15/2024 03:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____